

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 17 2009

Supervisor/Operator Performing the Verification Procedure:	
Name <u>BRANDON LAWRENSEN</u>	ID# <u>6609</u> Date <u>09/11/09</u>
A Agency <u>JUNEAU POLICE</u>	Phone # <u>586-0600</u>
Instrument Location <u>JUNEAU POLICE DEPT</u>	
Alco S/N <u>X124611</u> Target Value <u>.081</u> High Pressure <u>1200</u>	
B Alco Test Values	<u>.090</u> <u>.082</u> 1 st Alco 2 nd Alco
Signature <u>[Signature]</u> 008 9/28/09	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 2009.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130192

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

SEPTEMBER 11, 2009

OPERATOR'S NAME:

LAWRENSON/BRANDON/J

OPERATOR'S NUMBER: 6609

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0123456

DEPT/AGENCY: JNU1

CASE/REPORT: 0123456

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: X124611

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.30 in
ALCO TARGET .082 19:50
BLANK TEST .000 19:50
INTERNAL STANDARD VERIFIED 19:50
ALCO TV 30.30 in .090 19:51
BLANK TEST .000 19:52
SUBJECT SAMPLE .000 19:52
BLANK TEST .000 19:53
ALCO TV 30.30 in .082 19:53
BLANK TEST .000 19:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

SEPTEMBER 11, 2009
TIME 19:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.30 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~*~